

VISIA-CR

PRIMOS

- The outlet must provide power rated at least at 100-240VAC, 50/60Hz, 6 Amps.
- The lift workstation moves vertically to adjust for patient height. It is best to ensure that there is at least **(230cm)** from the floor available where the system will be positioned.

